

Quarter diopter IOL | Bi-Aspheric

Softec



High Definition Vision Preventing Negative Dysphotopsia

LENSTECO™ LENSTECO™ LENSTECO™ LENSTECO™ LENSTECO™ LENSTECO™



Single piece hydrophilic acrylic
Lenstec Softec HDO Oval

- » HDO Oval IOL has the largest 5.75 mm x 6.50 mm optic size in one-piece foldable hydrophilic lens
- » 30% More Coverage than Standard IOL
- » FDA Approved quality for the most accurate cataract surgery

**LENS
TEC**

more information on
www.lenstec.com



Technical Specifications

Optic Size	5.75 x 6.50 mm
Optic Type	Bi-aspheric
Length	12.50 mm
Haptic Style	Modified C
Angulation	0 Degrees
Positioning Holes	0
Construction	1 Piece
Optic Material	Acrylic (26% Water Content)
A/C Depth	5.22 mm
A-Constant(Contact Biometry)	118.00

A-Constant Optimized (Non-Contact Biometry)

Haigis	a0 = 1.546 a1 = 0.40 a2 = 0.10
SRK/T	A = 118.43
Holladay1	sf = 1.47
Holladay2	5.22
Hoffer Q	pACD = 5.22

Diopter Steps

Whole	+5.00 to +36.00
Half	+10.50 to +29.50
Quarter	+15.00 to +25.00

Expand your horizon Softec HD Oval Accuracy Without Artifact

30% More Coverage than Standard IOLs !


All cataract patients have some artifact or light-scattering phenomenon after IOL implantation, regardless of implant style, size, or material. Surgeons see this in many forms, characterized as either TCS (temporal crescent syndrome), dysphotopsia, edge glare, or other monikers. The patient describes the phenomenon at the one-day post-op visit, he is told it is "normal" and will go away, or that "you will get used to it". After three months or so, most patients do accept it, although it is generally still there.

Softec HD New Gold Standard - 0.11 D Tolerance - 3x More Precise! Optical Prescription Selection and Tolerance Example: eye requires 24.25D Prescription to achieve optimal vision.

Industry Standard IOL			Softec HD IOL		
Lens Prescription	Allowed Tolerance	Max. Variance for a standard IOL	Lens Prescription	Allowed Tolerance	Max. Variance for a Softec HD IOL
24.00	±0.4	IOL in an eye that requires a 24.25D lens is 0.65D . (smaller number is better).	24.00	±0.11	IOL in an eye that requires a 24.25D lens is 0.11D . (smaller number is better).
24.50	±0.4		24.25	±0.11	
			24.50	±0.11	
			24.75	±0.11	

Softec HD is the only IOL designed to address both Spherical Aberration and Defocus. Defocus is a more significant aberration than Spherical Aberration.

Bi-Aspheric Equal Conic Zero aberration IOL. Softec HD addresses the issue of spherical aberration inherent in conventional monofocal spherical IOLs by adjusting the optic with a patented design on both the anterior and posterior surfaces.



Studies have shown that Aspheric IOLs provide patients with significant optical benefits over traditional spherical surface IOLs.^{2,12,17}

Significant Outcomes. Mean refractive outcome was found to be closer to intended outcome, Depth of field was significantly improved, and Critical print size for fluent reading was smaller when compared to a standard monofocal IOL.⁷

Proven quality - FDA approved

Lenstec is one of eight companies in the world certified by FDA (Food and Drug Administration) for the sale of an intraocular lens in the U.S. market. All products have CE certificate, are approved by BSI (British Standards Institute) and are ISO quality system certified.

Stability of the biomaterial from which the intraocular Lenstec lenses are made, is long-term study proven and confirmed by millions of implanted lenses worldwide.

2. Thilos L, Hong X, Bradley A, Chang X. Statistical variation of aberration structure and image quality in a normal population of healthy eyes. J. Opt. Soc. Am A, Vol 19. No 21/Dec 2002 » 7. Craig J, Shah S, Wolffsohn J. Clinical evaluation of the Softec HD aberration-free aspheric intraocular lens. Submitted for publication. » 12. Sarver E. Theoretical optical performance of an equal conic intraocular lens and comparison to spherical and aspheric IOLs. AAO Presentation 2005 » 17. Nanavaty M, et al. Wavefront aberrations, depth of focus, and contrast sensitivity with aspheric and spherical intraocular lenses: fellow eye study. J Cataract Refract Surg. 2009; 35: 663 - 671



Lenstec is an ISO 13485 Registered company manufacturing CE Marked products.

Lenstec UK

Lenstec Barbados Inc.
Lenstec House
Unit 8, Mariner Court Calder Park
Wakefield WF4 3FL, UK

Tel: +44 (0)1924 382 678
Fax: +44 (0)1924 850 454
Email: lenstecuk@lenstec.com

Lenstec Europe

Trenčianska 47
821 09 Bratislava
Slovak Republic

cell: +421 905 798 760
e-mail: jkriska@lenstec.com

World Headquarters

Lenstec, Inc.
1765 Commerce Ave. N.
St. Petersburg
FL 33716, USA

Tel: 727-571-2272
Fax: 727-571-1792
Email: lenstec@lenstec.com

Lenstec Barbados

Lenstec Barbados Inc.
Airport Commercial Centre
Pilgrim Road, Christ Church
BB1 7092, BARBADOS

Tel: 246-420-6795
Fax: 246-420-6797
Email: lenstecbarbados@lenstec.com